



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/525,245

02/22/2005

Hideharu Tsuchiya

122572

5379

25944 7590 02/03/2009
OLIFF & BERRIDGE, PLC
P.O. BOX 320850
ALEXANDRIA, VA 22320-4850

EXAMINER

HENKEL, DANIELLE B

ART UNIT

PAPER NUMBER

1797

MAIL DATE

DELIVERY MODE

02/03/2009

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/525,245	Applicant(s) TSUCHIYA, HIDEHARU	
	Examiner DANIELLE HENKEL	Art Unit 1797	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 24 October 2008.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 22,23,27-29,31,36,37 and 41-53 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 22-23, 27-29, 31, 36-37, 41-53 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☒ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date <u>2/22/05</u> . | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Summary

1. The amendment filed October 24, 2008 has been entered and fully considered.
2. Claims 22, 23, 27-29, 31, 36, 37, and 41-53 remain pending in the application.

Claim Rejections - 35 USC § 103

3. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

4. The factual inquiries set forth in *Graham v. John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:
 1. Determining the scope and contents of the prior art.
 2. Ascertaining the differences between the prior art and the claims at issue.
 3. Resolving the level of ordinary skill in the pertinent art.
 4. Considering objective evidence present in the application indicating obviousness or nonobviousness.
5. Claims 22, 23, 27-29, 37, 42, 47, and 48 are rejected under 35 U.S.C. 103(a) as being unpatentable over BAKER (US 4301252) in view of KITAGAWA (US 4629862).
 - a. With respect to claim 22, BAKER teaches an upwardly extending cylindrical outer wall (12) on a base platform (11) with an observation hole (17) centrally of low inner wall (13) (at central portion) with a countersink (18)

Art Unit: 1797

(container accommodating portion) that permits a conventional culture dish (19) (specimen container) (Column 3, lines 9-19, Figure 1). BAKER teaches moisture soaked annular pads (15, 16) (water reservoir, water tank) are located in the channel between the outer and inner walls surrounding the culture dish (Column 3, lines 14-16, Figure 1). The incubator taught by BAKER also includes a transparent, cylindrical cover plate (25) (lid) that seals off the upper opening of the incubator (Column 3, lines 26-28, Figure 2). BAKER also teaches a thermal heating means (39) (heater) to provide suitable survival temperature in the incubator (Column 3, lines 57-68, Figure 1). The incubator of BAKER also includes a hollow fiber tube (30) (gas pipe) connected to a supply tube (32) (gas supply tube) that is connected externally to a supply hose (33) to feed gases into the incubator (Column 3, lines 36-47, Figure 1). The observation hole (17) (light transmitting portion of unit) located in the center of bottom platform (11) and the transparent cover plate (25) (light transmitting portion of lid) allow continuous observation of cell cultures using either an upright or inverted microscope (transmitting light rays upwardly or downwardly) (Column 2, lines 17-23). BAKER teaches an observation hole located in the center of bottom platform and a transparent (light ray transmitting) cover plate (lid) allow continuous observation of cell cultures using either an upright or inverted microscope (Column 2, lines 17-23). BAKER does not explicitly disclose the heater to be a plate type, having laminate plates, an upper plate or frame. However, KITAGAWA teaches a sample heater for use in microscopes in which an embodiment includes a heater

Art Unit: 1797

plate with a through-hole (light ray transmitting portion) that allows an objective lens of a microscope to be inserted into (corresponding to unit and lid) (Column 7, lines 65-66) in which a culture container is placed on the heater plate (heating container from bottom) (Column 8, lines 1-2). KITAGAWA also discloses the heater is a resistor (heating element) laminated in entirety (lamine upper and lower plates) (Column 11, lines 19-24). KITAGAWA also teaches an embodiment in which a top plate (inter-seat) is disposed above the upper plate with a space (see Figure 8) from the upper plate and a frame (rest with seat) for supporting the laminate and top plate (Column 10, lines 5-22, Figure 8). At the time of the invention, it would have been obvious to one of ordinary skill in the art to modify the incubator for light microscopy of BAKER to include the laminate plate heater and frame as taught by KITAGAWA. The motivation would have been that KITAGAWA discloses the plate heater has a uniform distribution of temperature, increased thermal efficiency and reduced response time (Column 2, lines 54-61) and the laminate plate heater reduces manufacturing cost (Column 11, lines 27-29) while the frame allows the top plate and heater to be removable so they are used only when necessary (Column 10, lines 58-62).

b. With respect to claim 23, KITAGAWA teaches the top plate is rested on an inner flange (seat) of the frame (Column 10, lines 11-15).

c. With respect to claim 42, KITAGAWA teaches a sample heater for use in microscopes in which an embodiment includes a heater plate with a through-hole (light ray transmitting portion) that allows an objective lens of a microscope to be

Art Unit: 1797

inserted into (corresponding to unit and lid) (Column 7, lines 65-66) in which a culture container is placed on the heater plate (heating container from bottom) (Column 8, lines 1-2).

d. With respect to claim 47, BAKER does not explicitly disclose the incubator unit having a lid with slots for carrying out operations on the specimen within. However, KITAGAWA teaches the sample heater has a lid (4b) on top of the culture container (5) with a through hole for the lens of a microscope (10) located centrally (accommodating portion for specimen container) and a plurality of through holes (4d) (slots) with covers (11) for the entry of a reagent or insertion of a temperature sensor (operations carried out to the specimen) (Column 6, lines 45-57). KITAGAWA also teaches the lid is slidable (lid is adjusted to shift) on the upper edge of the wall in an airtight manner (closing the opening of the unit) (Column 6, lines 55-57). The sliding of the lid places the through holes in direct alignment with the sample container (See Figure 3). At the time of the invention, it would have been obvious to one of ordinary skill in the art to modify the incubator for light microscopy of BAKER to include the lid with through holes of KITAGAWA. The motivation would have been that KITAGAWA discloses the lid allows for the entry of reagents or temperature and humidity sensors while maintaining an airtight closure of the container (Column 6, lines 50-57).

e. With respect to claim 48, BAKER discloses a cover for the incubator comprising a cover plate (lid) with a viewing aperture centered above the culture dish (region of the accommodating portion) to allow the use of a high power

Art Unit: 1797

objective lens while retaining the ability to focus on the cell culture (Column 4, lines 37-34). BAKER does not explicitly disclose the aperture being covered with a cover plate that can be displaced. However, KITAGAWA teaches a lid for a sample heater with through holes which are open and closed by lids provided on the surface of the main lid (cover plate displaced relative to upper surface of lid) (Column 6, lines 48-52). At the time of the invention, it would have been obvious to one of ordinary skill in the art to modify the cover with an aperture in the incubator for light microscopy of BAKER to include the lids of KITAGAWA. The motivation would have been that KITAGAWA discloses the lid maintains an airtight closure of the container while still allowing the entry of necessary objects (Column 6, lines 50-57).

f. With respect to claim 27, BAKER teaches an upwardly extending cylindrical outer wall (12) on a base platform (11) with an observation hole (17) centrally of low inner wall (13) (at central portion) with a countersink (18) (container accommodating portion) that permits a conventional culture dish (19) (specimen container) (Column 3, lines 9-19, Figure 1). BAKER teaches moisture soaked annular pads (15, 16) (water reservoir, water tank) are located in the channel between the outer and inner walls surrounding the culture dish (Column 3, lines 14-16, Figure 1). BAKER also includes a transparent, cylindrical cover plate (25) (lid) that seals off the upper opening of the incubator (Column 3, lines 26-28, Figure 2). BAKER also teaches a thermal heating means (39) (heater) to provide suitable survival temperature in the incubator (Column 3, lines 57-68,

Art Unit: 1797

Figure 1). The incubator of BAKER also includes a hollow fiber tube (30) (gas pipe) connected to a supply tube (32) (gas supply tube) that is connected externally to a supply hose (33) to feed gases into the incubator (Column 3, lines 36-47, Figure 1). The observation hole (17) (light transmitting portion of unit) located in the center of bottom platform (11) and the transparent cover plate (25) (light transmitting portion of lid) allow continuous observation of cell cultures using either an upright or inverted microscope (transmitting light rays upwardly or downwardly) (Column 2, lines 17-23). BAKER teaches an observation hole located in the center of bottom platform and a transparent (light ray transmitting) cover plate (lid) allow continuous observation of cell cultures using either an upright or inverted microscope (Column 2, lines 17-23). BAKER does not explicitly disclose the unit adapted to be placed on the stage without contacting the heater. However, KITAGAWA teaches an embodiment of a sample heater for use in a microscope with a container that has an annular projection on the bottom (Column 12, lines 57-65). KITAGAWA shows the annular projection on the culture container prevents it from directly contacting the heater below (not to contact the plate heater) and interposes a plate seat between the heating plate and container (spacing between) (Figure 23). KITAGAWA further discloses that the container is mounted through the plate seat on the upper surface of the heater plate with the plate seat fit in the annular projection of the container (unit and heater are separable) (Column 12, lines 11-15). At the time of the invention, it would have been obvious to one of ordinary skill in the art to modify the

Art Unit: 1797

incubator for light microscopy of BAKER to include the plate heater not directly in contact with the culture container as taught by KITAGAWA because the spacing from the heater allows for a uniform distribution of temperature, increased thermal efficiency and reduced response time (Column 2, lines 54-61).

g. With respect to claim 28, the limitations of claim 27 are met by the above combination of BAKER and KITAGAWA. Additionally BAKER teaches providing the base with an outward extending portion (28) with holes (29) (fixtures) to provide means for fastening the incubator onto (on upper surface) suitable observation means (microscope stage) (Column 3, lines 31-35).

h. With respect to claim 37, the limitations of claims 27 and 28 are considered met by the above combination of BAKER and KITAGAWA. Additionally BAKER discloses a clip (centering member) biased toward the culture dish to hold it firmly in place in the countersink (tool fitting hole) to maintain optical alignment of the culture cells (center unit in hole) (Column 4, lines 8-12). BAKER also discloses portion of the platform extending outward (around the peripheral portion of the unit) which have holes to provide a means for fastening (in contact with fixtures) the incubator to the stage (Column 3, lines 31-35).

i. With respect to claim 29, BAKER teaches an upwardly extending cylindrical outer wall (12) on a base platform (11) with an observation hole (17) centrally of low inner wall (13) (at central portion) with a countersink (18) (container accommodating portion) that permits a conventional culture dish (19)

Art Unit: 1797

(specimen container) (Column 3, lines 9-19, Figure 1). BAKER teaches moisture soaked annular pads (15, 16) (water reservoir, water tank) are located in the channel between the outer and inner walls surrounding the culture dish (Column 3, lines 14-16, Figure 1). The incubator taught by BAKER also includes a transparent, cylindrical cover plate (25) (lid) that seals off the upper opening of the incubator (Column 3, lines 26-28, Figure 2). BAKER also teaches a thermal heating means (39) (heater) to provide suitable survival temperature in the incubator (Column 3, lines 57-68, Figure 1). The incubator of BAKER also includes a hollow fiber tube (30) (gas pipe) connected to a supply tube (32) (gas supply tube) that is connected externally to a supply hose (33) to feed gases into the incubator (Column 3, lines 36-47, Figure 1). The observation hole (17) (light transmitting portion of unit) located in the center of bottom platform (11) and the transparent cover plate (25) (light transmitting portion of lid) allow continuous observation of cell cultures using either an upright or inverted microscope (transmitting light rays upwardly or downwardly) (Column 2, lines 17-23). BAKER teaches an observation hole located in the center of bottom platform and a transparent (light ray transmitting) cover plate (lid) allow continuous observation of cell cultures using either an upright or inverted microscope (Column 2, lines 17-23). BAKER does not explicitly disclose a means for displacing the container relative to the unit. However, KITAGAWA teaches an embodiment of an incubator for a microscope in which the position of the specimen (culture) container in a heating unit (box) can be varied by displacing the container

Art Unit: 1797

horizontally relative to the unit on the accommodating portion (rest) from outside the unit (Column 7, lines 29-45, Figure 3). At the time of the invention it would have been obvious to one of ordinary skill in the art to modify the incubator of BAKER to include the positioning means as taught by KITAGAWA because it allows for a wide area of observation by changing the relative position of the sample without uncovering the unit so that the gas concentration and heated air inside are not lost and contaminants cannot enter (Column 4, lines 1-13).

6. Claims 31 and 46 are rejected under 35 U.S.C. 103(a) as being unpatentable over BAKER (US 4301252) in view of KITAGAWA (US 4629862) as applied to claims 22, 23, 27-29, 37, 42, 47, and 48 above, and further in view of KURESHY (US 5192506).
 - a. With respect to claim 31, BAKER teaches an upwardly extending cylindrical outer wall (12) on a base platform (11) with an observation hole (17) centrally of low inner wall (13) (at central portion) with a countersink (18) (container accommodating portion) that permits a conventional culture dish (19) (specimen container) (Column 3, lines 9-19, Figure 1). BAKER teaches moisture soaked annular pads (15, 16) (water reservoir, water tank) are located in the channel between the outer and inner walls surrounding the culture dish (Column 3, lines 14-16, Figure 1). The incubator taught by BAKER also includes a transparent, cylindrical cover plate (25) (lid) that seals off the upper opening of the incubator (Column 3, lines 26-28, Figure 2). BAKER also teaches a thermal heating means

Art Unit: 1797

(39) (heater) to provide suitable survival temperature in the incubator (Column 3, lines 57-68, Figure 1). The incubator of BAKER also includes a hollow fiber tube (30) (gas pipe) connected to a supply tube (32) (gas supply tube) that is connected externally to a supply hose (33) to feed gases into the incubator (Column 3, lines 36-47, Figure 1). The observation hole (17) (light transmitting portion of unit) located in the center of bottom platform (11) and the transparent cover plate (25) (light transmitting portion of lid) allow continuous observation of cell cultures using either an upright or inverted microscope (transmitting light rays upwardly or downwardly) (Column 2, lines 17-23). BAKER teaches an observation hole located in the center of bottom platform and a transparent (light ray transmitting) cover plate (lid) allow continuous observation of cell cultures using either an upright or inverted microscope (Column 2, lines 17-23). BAKER does not explicitly teach the heater fit into a hole. However, KITAGAWA teaches a heater that is adapted to be fit into a tool fitting hole such that an upper surface of the heater on which the container is placed is flush with the upper surface of a portion of the stage of the microscope (Column 10, lines 5-22). At the time of the invention it would have been obvious to one of ordinary skill in the art to modify the incubator of BAKER to include the heater placed into a tool fitting hole as taught by KITAGAWA because it allows for the use of large sizes of containers which will rest stably on the heater and microscope stage (Column 10, lines 50-54). The combination of BAKER and KITAGAWA does not disclose an entrance opening. However, KURESHY teaches an incubator port closure system that

Art Unit: 1797

provides a door or shutter (side closure member) for closing or blocking the injection port (entrance opening) in the outer wall of the incubator (wall of the unit) for the insertion and removal of assay cartridges (Column 12, lines 60-68).

At the time of the invention, it would have been obvious to one of ordinary skill in the art to modify the incubator of BAKER and KITAGAWA to include the incubator door with closure as taught by KURESHY because it would be desirable to block the port when the cartridges are not being inserted or extracted from the incubator as the port permits air flow between the inside and outside of the incubator and causes undesirable variation in the chamber temperature (Column 2, lines 42-48).

- b. With respect to claim 46, BAKER and KITAGAWA do not disclose the incubator having an entrance opening in the side wall with a closure member for closing and opening the entrance. However, KURESHY teaches an incubator port closure system that provides a door or shutter (side closure member) for closing or blocking the injection port (entrance opening) in the outer wall of the incubator (wall of the unit) for the insertion and removal of assay cartridges (Column 12, lines 60-68). At the time of the invention, it would have been obvious to one of ordinary skill in the art to modify the incubator of BAKER and KITAGAWA to include the incubator door with closure of KURESHY. The motivation would have been that KURESHY discloses that it would be desirable to block the port when the cartridges are not being inserted or extracted from the incubator as the port

Art Unit: 1797

permits air flow between the inside and outside of the incubator and causes undesirable variation in the chamber temperature (Column 2, lines 42-48).

7. Claim 53 is rejected under 35 U.S.C. 103(a) as being unpatentable over BAKER (US 4301252) in view of KITAGAWA (US 4629862) as applied to claims 22, 23, 27-29, 37, 42, 47, and 48 above, and further in view of KURESHY (US 5192506) as applied to claims 31 and 46 above, and further in view of CHAN (US (US 6056342)).
 - a. With respect to claim 53, BAKER, KITAGAWA, and KURESHY do not explicitly disclose tongs for use with the incubator. However, CHAN discloses multi-purpose tongs with a U shaped handle (connected at rear ends), two gripper arms crossed in the middle, two holding members (urging portions) and anti-slide abutments (pinching portions) between the crossed arms and the holding members to prevent further movement of the arms (Column 3, lines 26-40). The tongs are opened by a user pinching the handle inwards (elastically deformable material) causing the holding members to separate. At the time of the invention, it would have been obvious to one of ordinary skill in the art to modify the incubator of BAKER, KITAGAWA, and KURESHY to include the tongs of CHAN. The motivation would have been that CHAN discloses the tongs are of efficient and low cost construction and versatile for use holding small articles and are easily cleanable (Column 2, lines 17-25).

Art Unit: 1797

8. Claims 41 and 43-44 are rejected under 35 U.S.C. 103(a) as being unpatentable over BAKER (US 4301252) in view of KITAGAWA (US 4629862) as applied to claims 22, 23, 27-29, 37, 42, 47, and 48 above, and further in view of DILLER (US 5257128).
 - a. With respect to claim 41, BAKER and KITAGAWA do not explicitly disclose a means for supplying water into the reservoir. The means for language used in claim 20 invokes 35 U.S.C. 112 6th paragraph interpretation such that the means for consists of a pipe, supply tube, and infusion reservoir as stated in the specification (Page 15). However DILLER teaches a perfusion stage for a microscope in which the perfusion system supplies fluid such as water through a fluid inlet tube (supply tube) from a fluid reservoir (infusion reservoir) which is fed through an inlet (pipe). At the time of the invention, it would have been obvious to one of ordinary skill in the art to modify the incubator of BAKER and KITAGAWA to include the perfusion system of DILLER. The motivation would have been that DILLER discloses the fluid system allows for serial perfusion of the specimen and rapid exchange of the bathing fluid with a low time constant while maintaining the specimen in focus of the microscopic field of view (Column 5, lines 30-35).
 - b. With respect to claim 43, BAKER and KITAGAWA do not explicitly disclose a nutrient medium supply means. The means for language used in claim 24 invokes 35 U.S.C. 112 6th paragraph interpretation such that the means for consists of a delivery tube, withdrawal tube, and media tank as stated in the specification (Page 14). However, DILLER teaches a perfusion stage for a

Art Unit: 1797

microscope in which the perfusion system connected to the sample chamber (container) includes a fluid inlet tube (tube for delivery), a fluid outlet tube (tube for withdrawal), and a fluid reservoir (media tank) connected to the inlet tube (Column 12, lines 17-20). At the time of the invention, it would have been obvious to one of ordinary skill in the art to modify the incubator for light microscopy of BAKER and KITAGAWA to include the perfusion system of DILLER. The motivation would have been that DILLER discloses the fluid system allows for serial perfusion of the specimen and rapid exchange of the bathing medium with a low time constant while maintaining the specimen in focus of the microscopic field of view (Column 5, lines 30-35).

- c. With respect to claim 44, BAKER and KITAGAWA do not explicitly disclose a nutrient supplying means that enables the replenishment of the nutrient medium without removing the lid of the unit. However, the perfusion stage for a microscope shown by DILLER in Figure 2 and described above in modified BAKER, allows for the inlet of fluid (71) from a reservoir (82) to the sample chamber (14) without the removal of the cover (lid) (12).

- 9. Claim 50 is rejected under 35 U.S.C. 103(a) as being unpatentable over BAKER (US 4301252) in view of KITAGAWA (US 4629862) as applied to claims 22, 23, 27-29, 37, 42, 47, and 48 above, and further in view of DILLER (US 5257128) as applied to claims 41, and 43-44 above, and further in view of ARGENTIERI (US 5241415).

Art Unit: 1797

- a. With respect to claim 50, the combination of BAKER, KITAGAWA, and DILLER does not explicitly disclose a heater with a heating portion formed of a transparent conductive film. However, ARGENTIERI discloses a heated recording chamber in which the preferable heater is a heating element consisting of a grid of wires embedded within a clear mylar film (transparent conductive film) (Column 3, lines 3-13). This heating element is then placed on the bottom of the tissue recording chamber (container-placing portion). At the time of the invention, it would have been obvious to one of ordinary skill in the art to modify the incubator of BAKER, KITAGAWA, and DILLER to include the clear mylar film encased heating element of ARGENTIERI. The motivation would have been that ARGENTIERI discloses the heating element below the tissue recording chamber keeps the temperature constant of the liquid bath in the chamber (Column 3, lines 1-4). ARGENTIERI also discloses that the wires of the heater are in a grid which provides transparent windows so that the tissue and cells in the chamber are illuminated with the same intensity as if the wires of the heating element were not present (Column 3, lines 58-67).

10. Claim 45 is rejected under 35 U.S.C. 103(a) as being unpatentable over BAKER (US 4301252) in view of KITAGAWA (US 4629862) as applied to claims 22, 23, 27-29, 37, 42, 47, and 48 above, and further in view of FOCHT (US 5552321).

- a. With respect to claim 45, BAKER discloses a clip to hold the culture dish in the countersink to maintain alignment of the culture cells (Column 4, lines 8-11).

Art Unit: 1797

Baker and KITAGAWA do not explicitly disclose the use of a pair of holders across the central portion of the unit. However, FOCHT discloses a culture dish apparatus that has retaining tabs (Figure 1, #13) (container holders) that are pivoted or slideably (adjustable spacing) moved to hold the culture dish in the recess (container accommodating portion) located in the center of the stage insert (Figure 1, Column 3, lines 65-67). At the time of the invention, it would have been obvious to one of ordinary skill in the art to modify the incubator of BAKER and KITAGAWA to include the retaining tabs of FOCHT. The motivation would have been the retaining tabs lock the culture dish assembly into position to allow proper connection to the heating elements of the incubator as FOCHT discloses (Column 4, lines 1-12).

11. Claims 49 and 52 are rejected under 35 U.S.C. 103(a) as being unpatentable over BAKER (US 4301252) in view of KITAGAWA (US 4629862) as applied to claims 22, 23, 27-29, 37, 42, 47, and 48 above, and further in view of ARGENTIERI (US 5241415).

- a. With respect to claim 49, BAKER and KITAGAWA do not explicitly disclose the bottom surface of the water tank is provided with a water tank heater. However, ARGENTIERI teaches a heated recording chamber in which a heater is adhered to the bottom of a serpentine passageway (water tank) to heat the liquid flowing into the cavity containing the cell or tissue sample (Column 3, lines 47-52). At the time of the invention, it would have been obvious to one of ordinary skill in the

Art Unit: 1797

art to modify the incubator of BAKER and KITAGAWA to include the fluid heater of ARGENTIERI. The motivation would have been that ARGENTIERI discloses exposing the fluid to the heater allows the temperature of the liquid to reach and stabilize at the temperature required of the liquid bath in the cavity (Column 3, lines 52-56).

- b. With respect to claim 52, BAKER discloses a clip to hold the culture dish in the countersink to maintain alignment of the culture cells (Column 4, lines 8-11). BAKER and KITAGAWA do not explicitly disclose the means for securing the specimen container to be the holders with screws as defined in the specification according to 35 U.S.C. 112 6th paragraph interpretation of the claim. However ARGENTIERI teaches a heated recording chamber where the tissue recording chamber (specimen container) is retained in position on a microscope stage through the use of clamps (holders) secured to the stage by knurled screws (Column 2, lines 40-49). At the time of the invention, it would have been obvious to one of ordinary skill in the art to modify the incubator of BAKER and KITAGAWA to include the clamps and screws of ARGENTIERI. The motivation would have been that ARGENTIERI discloses the clamps allow relatively precise positioning of the chamber with respect to the objective lens which must be brought into close proximity to the sample or cell in the recording chamber (Column 2, lines 49-55).

Art Unit: 1797

12. Claim 51 is rejected under 35 U.S.C. 103(a) as being unpatentable over BAKER (US 4301252) in view of KITAGAWA (US 4629862) as applied to claims 22, 23, 27-29, 37, 42, 47, and 48 above, and further in view of MIDDLEBROOK (US 5181382).

- a. With respect to claim 51, BAKER discloses the incubator with a culture dish (body open) with an upper surface (lid) (Column 4, line 40). Neither BAKER nor KITAGAWA explicitly disclose a lid provided with connections for the tubes. However, MIDDLEBROOK teaches a stage assembly in which the upper plate (lid) has three holes (aperture for connecting tube, channel) that extend angularly downwardly and centrally inwardly into the specimen chamber (lower surface of lid) (Column 6, lines 53-60). These holes are then connected to perfusion tubing, aspiration tubing or a sensor (Column 6, lines 57-64). At the time of the invention, it would have been obvious to one of ordinary skill in the art to modify the incubator of BAKER and KITAGAWA to include the lid with tube connections of MIDDLEBROOK. The motivation would have been that the system of MIDDLEBROOK does not require bulky positioning devices that compete for available space on the microscope stage (Column 1 line 66- Column 2 line 4).

13. Claim 36 is rejected under 35 U.S.C. 103(a) as being unpatentable over BAKER (US 4301252) in view of KITAGAWA (US 4629862) as applied to claims 22, 23, 27-29, 37, 42, 47, and 48 above, and further in view of INOUE (US 5717190)

- a. With respect to claim 36, BAKER teaches an upwardly extending cylindrical outer wall (12) on a base platform (11) with an observation hole (17) centrally of low

Art Unit: 1797

inner wall (13) (at central portion) with a countersink (18) (container accommodating portion) that permits a conventional culture dish (19) (specimen container) (Column 3, lines 9-19, Figure 1). BAKER teaches moisture soaked annular pads (15, 16) (water reservoir, water tank) are located in the channel between the outer and inner walls surrounding the culture dish (Column 3, lines 14-16, Figure 1). The incubator taught by BAKER also includes a transparent, cylindrical cover plate (25) (lid) that seals off the upper opening of the incubator (Column 3, lines 26-28, Figure 2). BAKER also teaches a thermal heating means (39) (heater) to provide suitable survival temperature in the incubator (Column 3, lines 57-68, Figure 1). The incubator of BAKER also includes a hollow fiber tube (30) (gas pipe) connected to a supply tube (32) (gas supply tube) that is connected externally to a supply hose (33) to feed gases into the incubator (Column 3, lines 36-47, Figure 1). The observation hole (17) (light transmitting portion of unit) located in the center of bottom platform (11) and the transparent cover plate (25) (light transmitting portion of lid) allow continuous observation of cell cultures using either an upright or inverted microscope (transmitting light rays upwardly or downwardly) (Column 2, lines 17-23). BAKER teaches an observation hole located in the center of bottom platform and a transparent (light ray transmitting) cover plate (lid) allow continuous observation of cell cultures using either an upright or inverted microscope (Column 2, lines 17-23). BAKER does not explicitly disclose the lid having a heating portion. However, KITAGAWA discloses a sample heater for use in microscopes in which the

Art Unit: 1797

container is heated from two opposite sides, including a sheet heater on the top (lid of incubator) of the culture container (Column 11, line 67- Column 12, line 3). At the time of the invention, it would have been obvious to one of ordinary skill in the art to modify the incubator of BAKER to include the lid heater as taught by KITAGAWA because the addition of a heater on the top plate allows for the container to be rapidly heated preventing the formation of water drops on the upper inner plate of the container which reduce visibility in the container (Column 12, lines 1-10). The combination of BAKER and KITAGAWA does not explicitly disclose the lid heater to have a conductive film layer disposed between glass plates. However, INOUE teaches a transparent heater for microscopes which has a laminate structure including a transparent conductive film layer disposed between transparent glass plates (Column 12, line 22 – Column 13, line 16). At the time of the invention it would have been obvious to modify the lid heater of BAKER and KITAGAWA to include the laminate structure with glass plates as taught by INOUE because it can heat the vicinity of the center of the heater used for examination at a uniform temperature without needing a through hole (Column 2, lines 16-22) and the glass plates have high transparency and low heat conductivity (Column 12, lines 63-64).

Response to Arguments

14. The previous objections to the disclosure for informalities are withdrawn in light of Applicant's amendments to the specification.

Art Unit: 1797

15. The previous objection to the information disclosure statement filed February 22, 2005 has been withdrawn in light of Applicant's arguments, and all information referred to therein has been considered (see updated PTO-1449 form).
16. The previous objections to claims 23 and 33 are withdrawn in light of Applicant's amendment to claim 23 and cancellation of claim 33.
17. The previous 35 U.S.C. 112 rejection of claim 27 is withdrawn in light of Applicant's amendment to claim 27.
18. Applicant's arguments, see pages 12-13, filed October 24, 2008, with respect to the rejection of claims 22 and 31 under 35 USC 103(a) have been fully considered and are persuasive. Therefore the rejections have been withdrawn. However, upon further consideration, new grounds of rejection are made in view of KITAGAWA (US 4629862).
19. In response to Applicant's argument traversing the rejection of claim 27, has been considered and is not persuasive. Applicant's argument that BAKER does not teach spacing between the water tank unit and heater does not take into account the combination with KITAGAWA which teaches a container spaced apart from the heater. KITAGAWA teaches that the container may have a water tank (water vessel) surrounding the culture sample (Column 7, lines 5-18). KITAGAWA also teaches that the embodiment shown in Figure 23 where the container is shown as a culture container is a representation and that the container shown is not limited to a laboratory dish culture container, but could be any container with the described projections that form the spacing (Column 14, lines 1-4). Therefore it would have

Art Unit: 1797

been obvious that the aforementioned water tank container of KITAGAWA could have the projections to achieve the same spacing. Therefore the combination of BAKER and KITAGAWA does teach all limitations of claim 27.

20. Applicant's arguments, see page 11, filed October 24, 2008, with respect to the rejection of claim 29 under 35 USC 103(a) have been fully considered and are persuasive. Therefore the rejection has been withdrawn. However, upon further consideration, a new ground of rejection is made in view of KITAGAWA (US 4629862).

21. Applicant's arguments, see page 14, filed October 24, 2008, with respect to the rejection of claim 36 under 35 USC 103(a) have been fully considered and are persuasive. Therefore the rejection has been withdrawn. However, upon further consideration, a new ground of rejection is made in view of INOUE (US 5717190).

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to DANIELLE HENKEL whose telephone number is (571)270-5505. The examiner can normally be reached on Mon-Thur: 7:30am-5pm, Alternate Fridays: 7:30am-4.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Jill Warden can be reached on 571-272-1267. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 1797

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/William H. Beisner/
Primary Examiner, Art Unit 1797

/DANIELLE HENKEL/
Examiner, Art Unit 1797